

Declaration of Conformity
according to the
Medical Devices Directive, 93/42/EEC
and
RoHS directive, 2011/65/EU

STANDARDS TO WHICH CONFORMITY IS DECLARED:

EN 60 601-1:2005

EN 50581:2012

EN 60 601-1-2

**Including: EN 61000-3-2, EN 61000-3-3, EN 55011 KI B, EN 61000-4-2, EN 61000-4-3, EN 61000-4-4,
EN 61000-4-5, EN 61000-4-6, EN 61000-4-8, EN 61000-4-11**

Manufacturer: Abelko Innovation
Address: 972 54 Luleå
Telephone: +46 920-220360
Telefax: +46 920-220068

Marketing: Ljungberg&Kögel AB
Address: Box 1032, 251 10 Helsingborg
Telephone: +46 042-139860
Telefax: +46 042-132181

Type of Equipment: Blood mixer
Model: BM330-1 and BM323-1
Product class: Class 1

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive and Standards.

Date of issue: 14 Apr 2014

Signature: 

/ Thommy Lundström

Position/title: CEO

